

DBH
pd

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/9/2015

Instrument Serial Number: 010568

As of 09-Apr-15

Start Date	End Date	Agency Name
08-Apr-15		DFS Central Lab
22-Oct-14	08-Apr-15	Mathews County SO
11-Jul-14	22-Oct-14	DFS Central Lab
28-May-14	11-Jul-14	Amherst County SO
22-Nov-13	28-May-14	DFS Central Lab
29-Apr-13	22-Nov-13	Prince William County PD
12-Mar-13	29-Apr-13	DFS Central Lab
21-Nov-11	12-Mar-13	Richmond PD
17-Nov-11	21-Nov-11	DFS Central Lab
07-Sep-11	17-Nov-11	Richmond PD
12-Feb-10	07-Sep-11	DFS Central Lab
16-Apr-09	12-Feb-10	Virginia Beach PD
23-Jun-08	16-Apr-09	DFS Central Lab

DBH
Jed

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010568** Worksheet Start Date **4/8/2015**
Location **Mathews CO SO**
Address **10622 Buckley Hall Rd Mathews VA 23109**
DFS Technician **Donald Hall** License No. **21250**

☐ Laboratory

☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) **764** Reference Barometer (mm HG) **766**
Reference Barometer(RB)Serial # **009109** RB Calibration Due **7/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.307
0.300	0.302	0.292	0.311	Sample 2	0.306
Precision		sample min	sample max	Sample 3	0.306
0.001		0.306	0.307		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.101
0.100	0.101	0.097	0.104	Sample 2	0.101
Precision		sample min	sample max	Sample 3	0.101
0		0.101	0.101		

Dry gas standard Lot No. (with tank no.) **AG320502-29**

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

Notes:

Instrument Serial Number

010568

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

4/9/15


Issuing Analyst



Date

4/13/15

DBH
fnd




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 04/08/2015
TEST EQUIPMENT NUMBER 010568		

RESULTS: TIME SAMPLE TAKEN 13:09 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 09 Apr 2015 at 09:15

DBH

gud

Test Results

Instrument Serial Number 010568

Test # 002114 Subject Test

Test Location 1 Department of
Test Date 08 Apr 2015

Test Location 2 Forensic Science
Test Time 13:02
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B
License Number 21250

Card Serial Number 121250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 08 Apr 2015 End Time 13:10

Result Time 13:09

Result Date 08 Apr 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 13:03

Data Type BLK

Sample Value 0.000

Sample Time 13:03

Data Type CHK

Sample Value 0.101

Sample Time 13:04

Data Type BLK

Sample Value 0.000

Sample Time 13:05

Data Type SUBJ

Sample Value 0.000

Sample Time 13:06

Data Type BLK

Sample Value 0.000

Sample Time 13:07

Data Type SUBJ

Sample Value 0.000

Sample Time 13:09

Data Type BLK

Sample Value 0.000

Sample Time 13:10

Standard Type Dry Gas Std

Standard Value 0.101

Standard Lot Number AG320502-29

Standard Expiration Date 07/24/2015

Tank Pressure 958

Barometric Pressure 764 mmHg

Blow Sample Number 1 Blow Duration 3.75 sec

Blow Volume 2040 cc End-of-Blow Time 13:06

Blow Sample Number 2 Blow Duration 3.31 sec

Blow Volume 1747 cc End-of-Blow Time 13:09

Tamper Evident Stamp c4e8a69f

Test Status Code 0

Test Status Success

DBH
Jed

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010568 Test Number: 2115
Test Date: 04/08/2015 Test Time: 13:12 EDT
Dry Gas Target: 0.101
Lot Number: AG320502-29 Exp Date: 07/24/2015
Tank Pressure: 953 psi Barometric Pressure: 764 mmHg
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:12
CHK	0.101	13:13
BLK	0.000	13:15
CHK	0.101	13:16
BLK	0.000	13:17
CHK	0.101	13:18

Test Status: *Success*

Calibration CRC: DF6B720B

DB/H
fnd

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010568 Test Number: 2116

Test Date: 04/08/2015 Test Time: 13:20 EDT

Dry Gas Target: 0.302

Lot Number: AG428003-05 Exp Date: 10/07/2016

Tank Pressure: 335 psi Barometric Pressure: 764 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:20
CHK	0.307	13:21
BLK	0.000	13:23
CHK	0.306	13:23
BLK	0.000	13:25
CHK	0.306	13:26

Test Status: *Success*

Calibration CRC: DF6B720B